

Wide Band Solid State EMC-Benchtop Power Amplifier 8GHz~11GHz



Features

- Automatic Calibration
- Built in Temperature Compensation
- Adjustable Attenuation:
31.5dB Range, 0.5dB Step Size
- Supply Voltage: 110V/220V AC

Typical Applications

- Aerospace and military applications
- Test and Measurement.
- Research and Development.

Electrical Specifications, $T_A=25\text{ }^{\circ}\text{C}$

Parameter	Min.	Typ.	Max.	Units
Frequency Range	8 - 11			GHz
Gain		65		dB
Gain Variation Over Temperature (-40°C~+85°C)		±3		dB
Input Return Loss		15		dB
Output Return Loss		20		dB
Saturated Output Power (Psat)		53		dBm
Isolation S12		-		dB
Input Max Power (No Damage)	Psat – Gain			dBm
Weight	560			ounces
Impedance	50			Ohms
Power Supply Connector	D-SUB COMBO 3POS			
Input / Output Connectors	Input SMA-Female or N-type Female --- Output N-Type Female			
Material	Aluminum / Copper			

Wide Band Solid State EMC-Benchtop Power Amplifier 8-11GHz

Absolute Maximum Ratings

Supply Voltage	110V 220V AC
RF Input Power (RFIN) Pin_max = Psat - Gainsat	Psat – Gain

Biasing Up Procedure

Step 1	Connect input and output with 50 Ohm source/load. (in band VSWR<1.9:1 or >10dB return loss)
Step 2	Turn on AC power.
Step 3	Enable RF output
Power OFF Procedure	
Step 1	Turn off RF output power
Step 2	Turn Off AC power
Step 3	Disconnect input and output

Environmental Specifications and Test Standards

Parameter	Description
Operational Temperature	-40°C~+85°C (Case Temperature)
Storage Temperature	-50°C~+105°C
Thermal Shock	-40°C → +85°C (5 Cycles / 10 hours)
Random Vibration	MIL-STD-202G Table 214-I, Test Condition Letter C 1.5 Hours Per Axis
High Temperature Burn In	Temperature +85°C for 72 Hours
Shock	1. Weight >20g, 50g half sine wave for 11ms, Speed variation 3.44m/s 2. Weight <=20g, 100g Half sine wave for 6ms, Speed variation 3.75m/s 3. Total 18 times (6 directions, 3 repetitions per direction).
Altitude	Standard: 30,000 Ft (Epoxy Sealed Controlled Environment) Optional: Hermetically Sealed (60,000 ft. 1.0 PSI min)
Hermetically Sealed (Optional)	MIL-STD-883 (For Hermetically Sealed Units)

Ordering Information

Part No.	Description
REMC08G11GD	8GHz~11GHz Power Amplifier

Amplifier Use

Ensure that the amplifier input and output ports are safely terminated into a proper 50 ohm load before turning on the power. Never operate the amplifier without a load. A proper 50 ohm load is defined as a load with impedance less than 1.9:1 or return loss larger than 10dB relative to 50 Ohm within the specified operating band width.

Power Supply Requirements

Power supply must be able to provide adequate current for the amplifier. Power supply should be able to provide 1.5 times the typical current or 1.2 times the maximum current (whichever is greater).

In most cases, RF - Lambda amplifiers will withstand severe mismatches without damage. However, operation with poor loads is discouraged. If prolonged operation with poor or unknown loads is expected, an external device such as an isolator or circulator should be used to protect the amplifier.

Ensure that the power is off when connecting or disconnecting the input or output of the amp.

Prevent overdriving the amplifier. Do not exceed the recommended input power level.

Adequate heat-sinking required for RF amplifier modules. Please inquire.

Amplifiers do not contain Thermal protection, Reverse DC polarity or Over voltage protection with the exception of a few models. Please inquire.

Proper electrostatic discharge (ESD) precautions are recommended to avoid performance degradation or loss of functionality.


What is not covered with warranty?

Each RF - Lambda amplifier will go through power and temperature stress testing.


Since the die, ICs or MMICs are fragile, these are not covered by warranty. Any damage to these will NOT be free to repair.

Typical Performance Plots


Gain




Input Return Loss



Isolation





Output Return Loss




Note: Input/output return loss measurements include attenuators to protect equipment

Wide Band Solid State EMG-Benchtop Power Amplifier 8-11GHz


Gain vs. Output Power




Pout vs. Pin



Saturation Power vs. Frequency




PxdB vs. Frequency




EMC Equipment User Manual

Front Panel





Rear Panel




Front Panel LCD Screen Display

Switching On Instrument



Please follow the instructions on the front panel LCD screen after switching on the power.
Press "1" on keypad to continue.

Self Calibration Screen






Calibration is may be recommended
"[1] Calibrate" to execute instrument self calibration process.

"[2] Reset" to reboot the instrument.

*Please turn OFF RF input power, and terminate the RF output port while applying calibration function

Instrument Protection Alarms




The front panel LCD screen will display the error code or error message when instrument self protection is triggered.
Front panel alarm indicator will light up.

To eliminate the error code, press "RESET" on front panel keypad to reboot the instrument and clear the alarms.

If error code can not be eliminated after reboot, please contact support@rflambda.com

Instrument Status Display Page



RF-LAMBDA
THE LEADER OF RF BROADBAND SOLUTIONS

Caution:

Temperature: 85 / 29.4 F/C
 Attenuation: 15 dB
 Frequency: 27 GHz
 Output Power: 50 dBm

[1] RF ON/OFF [2] Gain Lock
*press menu for more options

Indicates instrument RF output status. It will display:
Output is Ready to Turn on or RF Output is ON

Instrument temperature

RF output attenuation (change with adjustment knob)

RF input signal center frequency


Instrument RF output power

Press "Menu" on keypad to enter instrument functions selection menu

User can set a constant gain for the unit. Equipment will automatically adjust the gain at certain frequency

Switches On or Off for instrument RF output port

Instrument Function Selection Page



RF-LAMBDA
THE LEADER OF RF BROADBAND SOLUTIONS

Caution:

[1] Calibrate	[2] Frequency
[3] RF ON/OFF	[4] Reset
[5] Status	[6] Product Info

To enter this function selection page, press "Menu" on front panel keypad while the instrument is showing the status page.
Press the corresponding number on front panel keypad to select:
"[1] Calibrate" calibrates the instruments.


"[2] Frequency" enters RF input signal center frequency.

"[3] RF ON/OFF" switches the RF output port on or off.

"[4] Reset" Restarts the instrument (Turns RF output off)

"[5] Status" enters instrument status display page.

"[6] Product Info" displays product part number and serial number




RF-LAMBDA
THE LEADER OF RF BROADBAND SOLUTIONS

Attention

 Excute Calibration?

[1] YES [2] NO




RF-LAMBDA
THE LEADER OF RF BROADBAND SOLUTIONS

Attention

 Excute Reset?

[1] YES [2] NO



RF-LAMBDA
THE LEADER OF RF BROADBAND SOLUTIONS


Attention

 Turn Off RF Output?

[1] YES [2] NO

All action functions will ask for confirming execution when selected from function selection menu.

User Control Connector on Rear Panel




Pin #	Name	Function	Initial State	Description	Applied
1	Reset	Control		Resets PA when logic LOW is applied and released	Yes
2	Driver Disable	Control	LOW	Applying logic HIGH disables driver of amplifiers	Yes
3	Drain Disable	Control	LOW	Applying logic HIGH disables drain of amplifiers	Yes
4	RF IN Over	Indicator	LOW	Pin will be latched to logic HIGH when input signal is over limit	No
5	Temp Over	Indicator	LOW	Pin will be latched to logic HIGH when amplifier is driven over temperature	Yes
6	Current Over	Indicator	LOW	Pin will be latched to logic HIGH when drain current limit is reached	Yes
7	ID Imbalance	Indicator	LOW	Pin will be latched to logic HIGH when an imbalance in the drain current of the combining branches occurs	Yes
8	PA input power	Indicator		PA input power is represented by voltage	No
9	PA output power	Indicator		PA output power is represented by voltage	No
10	PA output reflection power	Indicator		PA output reflection power is represented by voltage	No
11	VSWR	Indicator	LOW	Pin will be latched to logic HIGH when output reflection is over limit	No
13	+5V	Power Supply	+5V	+5V DC is supplied for reference	Yes
14	GND	Ground	GND	Ground	Yes
15	GND	Ground	GND	Ground	Yes

HIGH/LOW voltages are standard TTL signals:
0.0V-0.8V = LOW
2V-5V = HIGH

Amplifier Outline Drawing:

All Dimensions in mm [inches]



Wide Band Solid State EMG-Benchtop Power Amplifier 8-11GHz

Important Notice

The information contained herein is believed to be reliable. RF-Lambda makes no warranties regarding the information contained herein. RF-Lambda assumes no responsibility or liability whatsoever for any of the information contained herein. RF-Lambda assumes no responsibility or liability whatsoever for the use of the information contained herein. The information contained herein is provided "AS IS, WHERE IS" and with all faults, and the entire risk associated with such information is entirely with the user. All information contained herein is subject to change without notice. Customers should obtain and verify the latest relevant information before placing orders for RF-Lambda products. The information contained herein or any use of such information does not grant, explicitly or implicitly, to any party any patent rights, licenses, or any other intellectual property rights, whether with regard to such information itself or anything described by such information.

RF-Lambda products are not warranted or authorized for use as critical components in medical, life-saving, or life sustaining applications, or other applications where a failure would reasonably be expected to cause severe personal injury or death.